

TE-1000 Hi-Vol Calibration Data Sheet

Site Information

Full Site Name: Xact - NIPSCO
 Site Abbreviation XAC Sampler Serial No. 49524619
 Field Technician Name Scott Keeler & Katie Healy Date 3/17/21 Time 9:35

Site Conditions

T_{amb} transfer standard (°C) 10.0 T_{amb} (K) 283.0 T_{amb} transfer standard (°C) + 273 = T_{amb} (K)
 P_{amb} transfer standard (mmHg) 744

Instructions: insert the temperature and pressure information into the TE-1000-PUF-Calibration-Worksheet.

Calibration Orifice Information

Make Graseby Model Regular Serial No. 584
 Orifice Slope "m_{orifice}" 10.46067 Orifice Intercept "b_{orifice}" -0.16706

Instructions: insert the orifice slope and intercept information into TE-1000-PUF-Calibration-Worksheet in the Q_{std} Slope and Q_{std} Intercept cells.

Hi-Vol Calibration Information

digital manometer

Test No.	Magnehelic Gauge Setting	High Manometer Pressure [†] (inH ₂ O)	Low Manometer Pressure [†] (inH ₂ O)	Sum of High & Low Pressures (inH ₂ O)	Q _{Manometer} (m ³ /min) from Tisch Calibration Worksheet	Q _{Magnehelic} (m ³ /min) from Tisch Calibration Worksheet
1	70.0	—	—	7.7	0.285	8.49
2	60.0	—	—	6.8	0.269	7.86
3	50.0	—	—	5.9	0.252	7.18
4	40.0	—	—	4.7	0.226	6.42
5	30.0	—	—	3.7	0.203	5.56

[†]If using a digital manometer, no need to fill out the High and Low columns, just the Sum column.

Instructions: (1) insert the data from "Sum of High & Low Pressures" column into the TE-1000-PUF-Calibration-Worksheet under the "Pressure (inH₂O)" column; (2) record the values from the "Q_{std} (m³/min)" column of the TE-1000-PUF-Calibration Worksheet under the "Q_{Manometer}" column in the table above; (3) record the values from the "Flow (corrected)" column of the TE-1000-PUF-Calibration Worksheet under the "Q_{Magnehelic}" column in the table above.

Hi-Vol Slope "m_{hivol}" 35.0115 Hi-Vol Intercept "b_{hivol}" -1.5447

Correction Coefficient "R" 0.9988 > 0.990? Yes No (circle one)

Instructions: record the values for from the "Slope", "Intercept", and "Corr. Coeff." cells found under "Linear Regression" in the TE-1000-PUF-Calibration Worksheet in the designated spaces above. If the Correction Coefficient is ≤ 0.990, repeat the calibration.

$$\text{Magnehelic Gauge Setpoint } 41.34 = \left(\frac{P_{amb}}{T_{amb}} * \frac{298K}{760\text{mmHg}} \right) * \left[\left(m_{hivol} * 0.225 \frac{\text{m}^3}{\text{min}} \right) + b_{hivol} \right]^2$$

Instructions: calculate the magnehelic gauge setpoint using the equation above.

P_{amb} = Atmospheric pressure from transfer standard, mmHg

T_{amb} = Atmospheric temperature from transfer standard, K